

ALAMO ANALYTICAL LABORATORIES, LTD.

Main: 10526 Gulfdale • San Antonio, Texas 78216-3601 • (210) 340-8121 . Fax. (210) 340-8123

REPORT NARRATIVE

5/26/2017

Chester Drash

Drash Consultants - TTL, Inc

1045 Central Parkway North, Suite 103

San Antonio , Texas - 78232

TEL: (210) 340-5004

Email: jmoheit@drashconsultants.com

FAX:

RE: 116E1164 1434 E. Commerce St

Dear Chester Drash:

Order No.: 1705108

Enclosed please find the analytical report for the sample/s received on 5/25/2017.

HOLDING TIMES: All samples were analyzed within prescribed holding times and/or in accordance with the Sample Acceptance Policy unless otherwise noted in the report.

DATA: Sample were prepared, analyzed and reported using the methods outlined in the following references: Test Methods for Evaluating Solid Waste, Physical/Chemical Methods, SW846, 3rd Edition

QA/QC: All method blanks, laboratory spikes, and/or matrix spikes met quality assurance objectives, except as noted in the report with data qualifiers.

SUBCONTRACTED: No analyses were subcontracted.

COMMENTS: Sample 1 and 3 diluted for TO-15 Analysis.

If you have any questions regarding these test results call (210) 340-8121.

Thank you,

Reddy Gosala, Ph.D

Laboratory Director

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17
Matrix: SOIL
Lab ID: 1705108-04A
Analyst: SLF

Client Sample ID B - 1 (8.5 - 10')

TOTAL PETROLEUM HYDROCARBONS TX1005

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
Hydrocarbons, C6-C12	< 50	20.2	50	1	mg/Kg	26-May-17	
Hydrocarbons, >C12-C28	< 50	24	50	1	mg/Kg	26-May-17	
Hydrocarbons, >C28-C35	< 50	24	50	1	mg/Kg	26-May-17	
Hydrocarbons, C6-C35	< 50	44.2	50	1	mg/Kg	26-May-17	

Surrogate Recoveries

Analyte	Recovery	Control Limits
1-Chlorooctadecane	79%	70—130
1-Chlorooctane	94%	70—130

For Surrogates: 0 = Dil. Out

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17
Matrix: SOIL
Lab ID: 1705108-05A
Analyst: SLF

Client Sample ID B - 2 (8.5 - 10')

TOTAL PETROLEUM HYDROCARBONS TX1005

Table with 7 columns: Analyte, Result, MDL, PQL, DF, Units, Date Analyzed. Rows include Hydrocarbons, C6-C12, >C12-C28, >C28-C35, and C6-C35.

Surrogate Recoveries

Table with 3 columns: Analyte, Recovery, Control Limits. Rows include 1-Chlorooctadecane and 1-Chlorooctane.

For Surrogates: 0 = Dil. Out

Approved by:

Handwritten signature

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.



Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17
Matrix: SOIL
Lab ID: 1705108-06A
Analyst: SLF

Client Sample ID B - 4 (2 - 3.5')

TOTAL PETROLEUM HYDROCARBONS TX1005

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
Hydrocarbons, C6-C12	< 50	20.2	50	1	mg/Kg	26-May-17	
Hydrocarbons, >C12-C28	< 50	24	50	1	mg/Kg	26-May-17	
Hydrocarbons, >C28-C35	< 50	24	50	1	mg/Kg	26-May-17	
Hydrocarbons, C6-C35	< 50	44.2	50	1	mg/Kg	26-May-17	

Surrogate Recoveries

Analyte	Recovery	Control Limits
1-Chlorooctadecane	108%	70—130
1-Chlorooctane	119%	70—130

For Surrogates: 0 = Dil. Out

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17
Matrix: AIR
Lab ID: 1705108-01A
Analyst: SS

Client Sample ID VW - 1

Volatiles by GC/MS method

TO-15

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
1,1,1-Trichloroethane	< 1	0.0168	1	2	ppbv	26-May-17	
1,1,2,2-Tetrachloroethane	< 1	0.0162	1	2	ppbv	26-May-17	
1,1,2-Trichloroethane	< 1	0.0132	1	2	ppbv	26-May-17	
1,1-Dichloroethane	< 1	0.0108	1	2	ppbv	26-May-17	
1,1-Dichloroethene	< 1	0.0158	1	2	ppbv	26-May-17	
1,2,4-Trichlorobenzene	< 1	0.03	1	2	ppbv	26-May-17	
1,2-Dibromo-3-chloropropane	< 1	0.03	1	2	ppbv	26-May-17	
1,2-Dibromoethane	< 1	0.032	1	2	ppbv	26-May-17	
1,2-Dichlorobenzene	< 1	0.026	1	2	ppbv	26-May-17	
1,2-Dichloroethane	< 1	0.0094	1	2	ppbv	26-May-17	
1,2-Dichloropropane	< 1	0.03	1	2	ppbv	26-May-17	
1,3-Dichlorobenzene	< 1	0.02	1	2	ppbv	26-May-17	
1,4-Dichlorobenzene	< 1	0.026	1	2	ppbv	26-May-17	
Benzene	18.5	0.0236	1	2	ppbv	26-May-17	
Bicycloheptadiene	< 1	0.03	1	2	ppbv	26-May-17	
Bromomethane	< 1	0.112	1	2	ppbv	26-May-17	
Carbon Tetrachloride	< 1	0.038	1	2	ppbv	26-May-17	
Chlorobenzene	< 1	0.022	1	2	ppbv	26-May-17	
Chloroform	< 1	0.032	1	2	ppbv	26-May-17	
Chloromethane	< 1	0.07	1	2	ppbv	26-May-17	
cis-1,2-Dichloroethene	< 1	0.0186	1	2	ppbv	26-May-17	
cis-1,3-Dichloropropene	< 1	0.036	1	2	ppbv	26-May-17	
Dichlorodifluoromethane	< 1	0.026	1	2	ppbv	26-May-17	
Dicyclopentadiene	< 1	0.012	1	2	ppbv	26-May-17	
Dimethyldisulfide	< 1	1.06	1	2	ppbv	26-May-17	
Ethylbenzene	44.4	0.0184	1	2	ppbv	26-May-17	
Furan	< 1	0	1	2	ppbv	26-May-17	
Hexachlorobutadiene	< 1	0.032	1	2	ppbv	26-May-17	
m,p-Xylene	171	6.33	150	300	ppbv	26-May-17	

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17

Client Sample ID VW - 1

Matrix: AIR
Lab ID: 1705108-01A
Analyst: SS

Volatiles by GC/MS method

TO-15

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
Methyl Isobutyl Ketone	< 1	0.126	1	2	ppbv	26-May-17	
Methyl tert butyl ether	< 1	0.0556	1	2	ppbv	26-May-17	
Methylene chloride	< 1	0.042	1	2	ppbv	26-May-17	
o-Xylene	13.2	0.0148	1	2	ppbv	26-May-17	
Styrene	< 1	0.016	1	2	ppbv	26-May-17	
Tetrachloroethene	< 1	0.018	1	2	ppbv	26-May-17	
Toluene	477	2.37	150	300	ppbv	26-May-17	
trans-1,2-Dichloroethene	< 1	0.0094	1	2	ppbv	26-May-17	
trans-1,3-Dichloropropene	< 1	0.046	1	2	ppbv	26-May-17	
Trichloroethene	< 1	0.0124	1	2	ppbv	26-May-17	
Trichlorofluoromethane	< 1	0.022	1	2	ppbv	26-May-17	
Vinyl chloride	< 1	0.0166	1	2	ppbv	26-May-17	
2-Butanone (MEK)	< 1	0.05	1	2	ppbv	26-May-17	
Tetrahydrofuran	< 1	0.018	1	2	ppbv	26-May-17	
1,1,2-Trichloro-1,2,2-trifluoroethane	< 1	0.24	1	2	ppbv	26-May-17	
1,2-Dichloro-1,1,2,2-tetrafluoroethane	< 1	0.09	1	2	ppbv	26-May-17	
Benzyl chloride	< 1	0.064	1	2	ppbv	26-May-17	
Bromochloromethane	< 1	0.026	1	2	ppbv	26-May-17	

Surrogate Recoveries

Analyte	Recovery	Control Limits
Bromofluorobenzene	107%	70—130

For Surrogates: 0 = Dil. Out

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17

Client Sample ID VW - 2

Matrix: AIR
Lab ID: 1705108-02A
Analyst: SS

Volatiles by GC/MS method

TO-15

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
1,1,1-Trichloroethane	< 10	0.168	10	20	ppbv	26-May-17	
1,1,2,2-Tetrachloroethane	< 10	0.162	10	20	ppbv	26-May-17	J
1,1,2-Trichloroethane	< 10	0.132	10	20	ppbv	26-May-17	
1,1-Dichloroethane	< 10	0.108	10	20	ppbv	26-May-17	
1,1-Dichloroethene	< 10	0.158	10	20	ppbv	26-May-17	
1,2,4-Trichlorobenzene	< 10	0.3	10	20	ppbv	26-May-17	
1,2-Dibromo-3-chloropropane	< 10	0.3	10	20	ppbv	26-May-17	
1,2-Dibromoethane	< 10	0.32	10	20	ppbv	26-May-17	
1,2-Dichlorobenzene	< 10	0.26	10	20	ppbv	26-May-17	
1,2-Dichloroethane	< 10	0.094	10	20	ppbv	26-May-17	
1,2-Dichloropropane	< 10	0.3	10	20	ppbv	26-May-17	
1,3-Dichlorobenzene	< 10	0.2	10	20	ppbv	26-May-17	
1,4-Dichlorobenzene	< 10	0.26	10	20	ppbv	26-May-17	
Benzene	< 10	0.236	10	20	ppbv	26-May-17	
Bicycloheptadiene	< 10	0.3	10	20	ppbv	26-May-17	
Bromomethane	< 10	1.12	10	20	ppbv	26-May-17	
Carbon Tetrachloride	< 10	0.38	10	20	ppbv	26-May-17	
Chlorobenzene	< 10	0.22	10	20	ppbv	26-May-17	
Chloroform	< 10	0.32	10	20	ppbv	26-May-17	
Chloromethane	< 10	0.7	10	20	ppbv	26-May-17	
cis-1,2-Dichloroethene	< 10	0.186	10	20	ppbv	26-May-17	
cis-1,3-Dichloropropene	< 10	0.36	10	20	ppbv	26-May-17	
Dichlorodifluoromethane	< 10	0.26	10	20	ppbv	26-May-17	
Dicyclopentadiene	< 10	0.12	10	20	ppbv	26-May-17	
Dimethyldisulfide	< 10	10.6	10	20	ppbv	26-May-17	
Ethylbenzene	< 10	0.184	10	20	ppbv	26-May-17	
Furan	< 10	0	10	20	ppbv	26-May-17	
Hexachlorobutadiene	< 10	0.32	10	20	ppbv	26-May-17	
m,p-Xylene	< 20	0.422	20	20	ppbv	26-May-17	J

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17
Matrix: AIR
Lab ID: 1705108-02A
Analyst: SS

Client Sample ID VW - 2

Volatiles by GC/MS method

TO-15

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
Methyl Isobutyl Ketone	< 10	1.26	10	20	ppbv	26-May-17	
Methyl tert butyl ether	< 10	0.556	10	20	ppbv	26-May-17	
Methylene chloride	< 10	0.42	10	20	ppbv	26-May-17	
o-Xylene	< 10	0.148	10	20	ppbv	26-May-17	
Styrene	74	0.16	10	20	ppbv	26-May-17	
Tetrachloroethene	< 10	0.18	10	20	ppbv	26-May-17	
Toluene	< 10	0.158	10	20	ppbv	26-May-17	J
trans-1,2-Dichloroethene	< 10	0.094	10	20	ppbv	26-May-17	
trans-1,3-Dichloropropene	< 10	0.46	10	20	ppbv	26-May-17	
Trichloroethene	< 10	0.124	10	20	ppbv	26-May-17	
Trichlorofluoromethane	< 10	0.22	10	20	ppbv	26-May-17	
Vinyl chloride	< 10	0.166	10	20	ppbv	26-May-17	
2-Butanone (MEK)	< 10	0.5	10	20	ppbv	26-May-17	
Tetrahydrofuran	< 10	0.18	10	20	ppbv	26-May-17	
1,1,2-Trichloro-1,2,2-trifluoroethane	< 10	2.4	10	20	ppbv	26-May-17	
1,2-Dichloro-1,1,2,2-tetrafluoroethane	< 10	0.9	10	20	ppbv	26-May-17	
Benzyl chloride	< 10	0.64	10	20	ppbv	26-May-17	
Bromochloromethane	< 10	0.26	10	20	ppbv	26-May-17	

Surrogate Recoveries

Analyte	Recovery	Control Limits
Bromofluorobenzene	117%	70—130

For Surrogates: 0 = Dil. Out

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17

Client Sample ID VW - 3

Matrix: AIR
Lab ID: 1705108-03A
Analyst: SS

Volatiles by GC/MS method

TO-15

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
1,1,1-Trichloroethane	< 10	0.168	10	20	ppbv	26-May-17	
1,1,2,2-Tetrachloroethane	< 10	0.162	10	20	ppbv	26-May-17	
1,1,2-Trichloroethane	< 10	0.132	10	20	ppbv	26-May-17	
1,1-Dichloroethane	< 10	0.108	10	20	ppbv	26-May-17	
1,1-Dichloroethene	< 10	0.158	10	20	ppbv	26-May-17	
1,2,4-Trichlorobenzene	< 10	0.3	10	20	ppbv	26-May-17	
1,2-Dibromo-3-chloropropane	< 10	0.3	10	20	ppbv	26-May-17	
1,2-Dibromoethane	< 10	0.32	10	20	ppbv	26-May-17	
1,2-Dichlorobenzene	< 10	0.26	10	20	ppbv	26-May-17	
1,2-Dichloroethane	< 10	0.094	10	20	ppbv	26-May-17	
1,2-Dichloropropane	< 10	0.3	10	20	ppbv	26-May-17	
1,3-Dichlorobenzene	< 10	0.2	10	20	ppbv	26-May-17	
1,4-Dichlorobenzene	< 10	0.26	10	20	ppbv	26-May-17	
Benzene	< 10	0.236	10	20	ppbv	26-May-17	
Bicycloheptadiene	< 10	0.3	10	20	ppbv	26-May-17	
Bromomethane	< 10	1.12	10	20	ppbv	26-May-17	
Carbon Tetrachloride	< 10	0.38	10	20	ppbv	26-May-17	
Chlorobenzene	< 10	0.22	10	20	ppbv	26-May-17	
Chloroform	< 10	0.32	10	20	ppbv	26-May-17	
Chloromethane	< 10	0.7	10	20	ppbv	26-May-17	
cis-1,2-Dichloroethene	< 10	0.186	10	20	ppbv	26-May-17	
cis-1,3-Dichloropropene	< 10	0.36	10	20	ppbv	26-May-17	
Dichlorodifluoromethane	< 10	0.26	10	20	ppbv	26-May-17	
Dicyclopentadiene	< 10	0.12	10	20	ppbv	26-May-17	
Dimethylsulfide	< 10	10.6	10	20	ppbv	26-May-17	
Ethylbenzene	< 10	0.184	10	20	ppbv	26-May-17	J
Furan	< 10	0	10	20	ppbv	26-May-17	
Hexachlorobutadiene	< 10	0.32	10	20	ppbv	26-May-17	
m,p-Xylene	20	0.422	20	20	ppbv	26-May-17	

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



ALAMO ANALYTICAL LABORATORIES, LTD.

Analytical Results Report

Client: Drash Consultants - TTL, Inc
Work Order: 1705108
Project Name: 116E1164 1434 E. Commerce St

Date: 26-May-17
Date Received: 25-May-17
Collection Date: 25-May-17
Preparation Date: 26-May-17

Client Sample ID VW - 3

Matrix: AIR
Lab ID: 1705108-03A
Analyst: SS

Volatiles by GC/MS method

TO-15

Analyte	Result	MDL	PQL	DF	Units	Date Analyzed	Qual
Methyl Isobutyl Ketone	< 10	1.26	10	20	ppbv	26-May-17	
Methyl tert butyl ether	< 10	0.556	10	20	ppbv	26-May-17	
Methylene chloride	< 10	0.42	10	20	ppbv	26-May-17	
o-Xylene	< 10	0.148	10	20	ppbv	26-May-17	J
Styrene	27.2	0.16	10	20	ppbv	26-May-17	
Tetrachloroethene	< 10	0.18	10	20	ppbv	26-May-17	
Toluene	1070	0.395	1	50	ppbv	26-May-17	
trans-1,2-Dichloroethene	< 10	0.094	10	20	ppbv	26-May-17	
trans-1,3-Dichloropropene	< 10	0.46	10	20	ppbv	26-May-17	
Trichloroethene	< 10	0.124	10	20	ppbv	26-May-17	
Trichlorofluoromethane	< 10	0.22	10	20	ppbv	26-May-17	
Vinyl chloride	< 10	0.166	10	20	ppbv	26-May-17	
2-Butanone (MEK)	< 10	0.5	10	20	ppbv	26-May-17	
Tetrahydrofuran	< 10	0.18	10	20	ppbv	26-May-17	
1,1,2-Trichloro-1,2,2-trifluoroethane	< 10	2.4	10	20	ppbv	26-May-17	
1,2-Dichloro-1,1,2,2-tetrafluoroethane	< 10	0.9	10	20	ppbv	26-May-17	
Benzyl chloride	< 10	0.64	10	20	ppbv	26-May-17	
Bromochloromethane	< 10	0.26	10	20	ppbv	26-May-17	

Surrogate Recoveries

Analyte	Recovery	Control Limits
Bromofluorobenzene	139%	70—130

For Surrogates: 0 = Dil. Out

Approved by:

Report of Laboratory Analysis

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.

NELAP Certificate# San Antonio : T104704367-17-7



CLIENT: Drash Consultants - TTL, Inc

Work Order: 1705108

Project: 116E1164 1434 E. Commerce St

QC SUMMARY REPORT

Analyte	%REC				%REC		RPD		Low - High		
	BLK	SPK value	LCS	LCSD	RPD %	RPD Limit	MS	MSD	%	Limit	Limit
Batch ID: AVOC_170526A	TestName: Volatiles by GC/MS method										
Run ID: AVOC_170526A	Test Code: TO-15		Units: ppbv			Analysis Date: 5/26/2017 12:45:00 PM			Prep Date: 5/26/2017 12:45:00		
1,1,1-Trichloroethane	<0.5	5	86.4%								70 - 130
1,1,2,2-Tetrachloroethane	<0.5	5	92.0%								70 - 130
1,1,2-Trichloroethane	<0.5	5	92.2%								70 - 130
1,1-Dichloroethane	<0.5	5	95.2%								70 - 130
1,1-Dichloroethene	<0.5	5	107.8%								70 - 130
1,2,4-Trichlorobenzene	<0.5	5	88.0%								70 - 130
1,2-Dibromo-3-chloropropane	<0.5	5	86.0%								70 - 130
1,2-Dibromoethane	<0.5	5	92.0%								70 - 130
1,2-Dichlorobenzene	<0.5	5	110.0%								70 - 130
1,2-Dichloroethane	<0.5	5	98.8%								70 - 130
1,2-Dichloropropane	<0.5	5	93.6%								70 - 130
1,3-Dichlorobenzene	<0.5	5	108.2%								70 - 130
1,4-Dichlorobenzene	<0.5	5	109.2%								70 - 130
Benzene	<0.5	5	99.2%								70 - 130
Bicycloheptadiene	<0.5	5	116.2%								70 - 130
Bromomethane	<0.5	5	105.4%								70 - 130
Carbon Tetrachloride	<0.5	5	88.8%								70 - 130
Chlorobenzene	<0.5	5	98.8%								70 - 130
Chloroform	<0.5	5	97.4%								70 - 130
Chloromethane	<0.5	5	95.0%								70 - 130
cis-1,2-Dichloroethene	<0.5	5	90.0%								70 - 130
cis-1,3-Dichloropropene	<0.5	5	92.2%								70 - 130
Dichlorodifluoromethane	<0.5	5	101.6%								70 - 130
Dicyclopentadiene	<0.5	5	82.4%								70 - 130
Dimethyldisulfide	<0.5	5	81.6%								70 - 130
Ethylbenzene	<0.5	5	80.8%								70 - 130
Furan	<0.5	5	79.8%								70 - 130
Hexachlorobutadiene	<0.5	5	84.6%								70 - 130
m,p-Xylene	<1	10	94.4%								70 - 130

Approved by: *[Signature]*

Laboratory QC Report

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.



CLIENT: Drash Consultants - TTL, Inc

Work Order: 1705108

Project: 116E1164 1434 E. Commerce St

QC SUMMARY REPORT

Analyte	BLK	SPK value	%REC				%REC		RPD		Low - High
			LCS	LCSD	RPD %	RPD Limit	MS	MSD	%	Limit	Limit
Methyl Isobutyl Ketone	<0.5	5	106.4%								70 - 130
Methyl tert butyl ether	<0.5	5	92.2%								70 - 130
Methylene chloride	<0.5	5	86.4%								70 - 130
o-Xylene	<0.5	5	106.8%								70 - 130
Styrene	<0.5	5	125.4%								70 - 130
Tetrachloroethene	<0.5	5	105.0%								70 - 130
Toluene	<0.5	5	80.2%								70 - 130
trans-1,2-Dichloroethene	<0.5	5	87.0%								70 - 130
trans-1,3-Dichloropropene	<0.5	5	98.0%								70 - 130
Trichloroethene	<0.5	5	90.4%								70 - 130
Trichlorofluoromethane	<0.5	5	107.2%								70 - 130
Vinyl chloride	<0.5	5	107.4%								70 - 130
2-Butanone (MEK)	<0.5	5	98.4%								50 - 150
Tetrahydrofuran	<0.5	5	90.0%								55 - 145
1,1,2-Trichloro-1,2,2-trifluoroethane	<0.5	5	115.8%								70 - 130
1,2-Dichloro-1,1,2,2-tetrafluoroethane	<0.5	5	115.8%								70 - 130
Benzyl chloride	<0.5	5	112.6%								70 - 130
Bromochloromethane	<0.5	5	95.2%								70 - 130

Batch ID: TPH1005_S-5/26/2017 TestName: TOTAL PETROLEUM HYDROCARBONS
Run ID: TPH_170526A Test Code: TX1005 Units: mg/Kg Analysis Date: 5/26/2017 Prep Date: 5/26/2017

Hydrocarbons, C6-C12	<50	500	100.8%	111.4%	10.000	30.0	84.6%	86.4%	2.000	30.0	75 - 125
Hydrocarbons, >C12-C28	<50	500	104.2%	103.8%	0.000	30.0	89.6%	84.0%	6.000	30.0	75 - 125
Hydrocarbons, >C28-C35	<50										
Hydrocarbons, C6-C35	<50	1000	102.5%	107.6%	6.000	30.0	87.1%	85.2%	2.000	30.0	75 - 125

Approved by: *[Signature]*

Laboratory QC Report

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.



CLIENT: Drash Consultants - TTL, Inc

Work Order: 1705108

Project: 116E1164 1434 E. Commerce St


QC SUMMARY REPORT

Analyte	%REC				%REC		RPD		Low - High	
	BLK	SPK value	LCS	LCSD	RPD %	RPD Limit	MS	MSD	%	Limit

Approved by: *[Signature]*

Laboratory QC Report

Note: The analysis contained in this report applies only to the samples tested and for the exclusive use of the addressed client. Reproduction of this report wholly or in part requires written permission of the client.



MUST BE COMPLETED BY CLIENT

Alamo's Client: Drash Consultants	Client's P.O. #:	Turnaround time: Standard(7) <input type="checkbox"/> (in working days) RUSH: 1 <input checked="" type="checkbox"/> 2 <input type="checkbox"/> 3-5 <input type="checkbox"/> Days (additional charges) TRRP 13 Report: Yes <input type="checkbox"/> No <input type="checkbox"/> (additional charges) Analysis for Permit Compliance: Yes <input type="checkbox"/> No <input type="checkbox"/> DMR Form Required: Yes <input type="checkbox"/> No <input type="checkbox"/>
Project Manager: Chester Drash	Phone #: (210) 888-6100	
Address: 1045 Central Pkwy N. Suite 103 jmoheit@drash	Email: consultants.com	
Project Number: 116E1164	Project Name:	
Project Location: 1434 E. Commerce St.	Sampler Signature: <i>Jeremy Moheit</i> Jeremy Moheit	

Main Office: 10526 Gulfdale
San Antonio, Texas 78216
(210) 340-8121 • Fax (210) 340-8123

Branch: 2500 Montana Avenue
El Paso, Texas 79903
(915) 599-2182

www.alamoanalytical.com
reports@alamoanalytical.com

LAB ID# (Do not use)	Sampling		Composite	Grab	Matrix	FIELD ID#	FIELD DESCRIPTION	No. of Containers	ANALYSIS				REMARKS (Preservation, Size/Amount, Etc.)
	Date	Time							VOCs	TPH			
1205108-21	5-25-17	1230		X	Air		VW-1	1	X				
-02		1245		X	↓		VW-2	1	X				
-03		1300		X	↓		VW-3	1	X				
-04		0820		X	soil		B-1 (8.5-10')	1		X			4oz Jar
-05		0930		X	↓		B-2 (8.5-10')	1		X			4oz Jar
-06		1110		X	↓		B-4 (2-3.5')	1		X			4oz Jar

Relinquished by: (Signature / Print Name) <i>Jeremy Moheit</i> / Jeremy Moheit	Date 5/25/17	Time 3:00 PM	Received by: (Signature)	Headspace <input type="checkbox"/> MA If Yes, Amt. _____ Properly Sealed <input checked="" type="checkbox"/> If No, Explain _____ Chilled ≤4° C <input checked="" type="checkbox"/> If No, Temp. _____ Comments:
Relinquished by: (Signature / Print Name)	Date	Time	Received by: (Signature)	
Relinquished by: (Signature / Print Name)	Date	Time	Received by: (Signature)	
Relinquished by: (Signature / Print Name)	Date 5/25/17	Time 3:00 PM	Received by: (Signature)	



ALAMO ANALYTICAL LABORATORIES, LTD.

10526 Gulfdale
San Antonio, TX - 78216;
Ph. (210) 340-8121; Fax: (210) 340-8123

www.alamoanalytical.com

2500 Montana Avenue
El Paso, TX - 79903
Ph. (915) 599-2182

Sample Log-In Checklist

DATE: 05/25/2017 TIME: 15:05 INITIALS: [Signature]

CLIENT: Rush PROJECT: W.O.# 1705108

1. Is a Chain of Custody present? Yes No
2. Is a Chain of Custody properly completed? Yes No
3. Are custody seals present? Yes No
If yes, are they intact? Yes No
 Are they on: Sample _____ of on Shipping Container _____
4. Are all samples tagged or labeled? Yes No
If yes, do the labels match the Chain of Custody? Yes No
5. Do all shipping documents agree (i.e., number of coolers arrived vs. on tickets)
If not, describe below. Yes No N/A
6. Are samples preserved properly? *If not, describe below.* Yes No
7. Are all samples within holding times on arrival?
If not, describe below. Yes No
8. Condition of shipping container: Intact or _____
9. Condition of samples: Intact or _____
10. Temperature of samples: Temp. (°C): Corrected Temp. (°C): Thermometer ID : DT1 or L2
11. pH strip lot#: Samples out of pH range:
12. Delivery agent: Client UPS Fed-Ex Lone Star Alamo P/U Other
13. Sample disposal: Return to client Alamo Analytical Disposal

Comments: (Reference checklist item number from above, or for comments on resolution below):
* 1705 108-04 to 06 Solids RUSH 24 hr on ice temp: 3.2°C
* 1705 108-01 to 03 Air RUSH 24 hr

Record of contacting client for resolution of sample discrepancies (first and retry contact) Contacted How?

Name: _____ Phone _____ Fax _____ Date: ____/____/____ Time: _____
Name: _____ Phone _____ Fax _____ Date: ____/____/____ Time: _____